



## Certificate of Analysis

**Product: SpermFilter<sup>®</sup> 100%**

Batch number: FLT100-807  
Expiration Date: 30-09-2019  
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,53
Osmolality:	300 – 330 mOsm/kg	310
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	< 0,1 EU/ml
Density:	1,1150 – 1,1250 g/ml	1,1177
Sperm Survival:	≥ 80%	99%

---

Approved on: 22-03-2018  
(DD/MM/YYYY)

By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to be 'M. Berbers', written over a horizontal line.

Quality Control Department, Gynotec B.V.