



Certificate of Analysis

Product: SpermFilter® 100%

Batch number: FLT100-814
Expiration Date: 30-11-2019
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,27
Osmolality:	300 – 330 mOsm/kg	303
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	0,118 EU/ml
Density:	1,1150 – 1,1250 g/ml	1,1201
Sperm Survival:	≥ 80%	99%

Approved on:
(DD/MM/YYYY)

By:

01-06-2018

J. Berbers

Signature:

Quality Control Department, Gynotec B.V.