



Certificate of Analysis

Product: SpermFilter® 100%

Batch number: FLT100-818
Expiration Date: 31-01-2020
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,28
Osmolality:	300 – 330 mOsm/kg	314
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	0,1 EU/ml
Density:	1,1150 – 1,1250 g/ml	1,1203
Sperm Survival:	≥ 80%	100%

Approved on: 06-08-2018
(DD/MM/YYYY)

By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to be 'M. Berbers', written over a horizontal line.

Quality Control Department, Gynotec B.V.