



## Certificate of Analysis

**Product: SpermFilter<sup>®</sup> Gradient 80%**

Batch number: FLT80-803  
Expiration Date: 31-07-2019  
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,49
Osmolality:	310 – 330 mOsm/kg	329
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	< 0,1 EU/ml
Density:	1,0970 – 1,1070 g/ml	1,0983
Sperm Survival:	≥ 80%	100%

---

Approved on: 13-02-2018  
(DD/MM/YYYY)

By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to be 'M. Berbers'.

Quality Control Department, Gynotec B.V.