



## Certificate of Analysis

**Product: SpermWash®**

Batch number: SW-804  
Expiration Date: 31-07-2019  
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,30 – 7,60	7,44
Osmolality:	270 – 290 mOsm/kg	282
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,25 EU/ml	< 0,05 EU/ml
MEA:	≥ 80%	95%

---

Approved on: 06-02-2018  
(DD/MM/YYYY)

By: M. Berbers

Signature:

Quality Control Department, Gynotec B.V.