



Certificate of Analysis

Product: SpermWash®

Batch number: SW-809
Expiration Date: 31-08-2019
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,30 – 7,60	7,41
Osmolality:	270 – 290 mOsm/kg	278
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,25 EU/ml	< 0,05 EU/ml
MEA:	≥ 80%	100%

Approved on: 23-02-2018
(DD/MM/YYYY)

By: M. Berbers

Signature:

Quality Control Department, Gynotec B.V.