



Certificate of Analysis

Product: SpermWash®

Batch number:

SW-822

Expiration Date:
(DD/MM/YYYY)

30-11-2019

Analysis	Limits	Results
pH:	7,30 – 7,60	7,44
Osmolality:	270 – 290 mOsm/kg	276
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,25 EU/ml	< 0,05 EU/ml
MEA:	≥ 80%	95%

Approved on:
(DD/MM/YYYY)

24-05-2018

By:

J. Berberg

Signature:

Quality Control Department, Gynotec B.V.